
GB55-12

## Peak Atlas DCA

Semiconductor Component Analyser
Model DCA55



Designed and manufactured with pride in the UK

## User Guide

 $\bigcirc$  Peak Electronic Design Limited 2000/2016 In the interests of development, information in this guide is subject to change without notice - E&OE





## Want to use it now?

We understand that you want to use your *Atlas DCA* right now. The unit is ready to go and you should have little need to refer to this user guide, but please make sure that you do at least take a look at the notices on page 4!

Contents	Page
Introduction	3
Important Considerations	4
Analysing semiconductors	5
Diodes	7
Diode Networks	8
LEDs	9
Bicolour LEDs	10
Bipolar Junction Transistors (BJTs)	11
Digital Transistors	18
Enhancement Mode MOSFETs	19
Depletion Mode MOSFETs	20
Junction FETs (JFETs)	
Thyristors (SCRs) and Triacs	22
Taking care of your <i>Atlas DCA</i>	23
Battery replacement	
Self Tests	24
Appendix A - Technical Specifications	
Appendix B - Warranty Information	
Annendix C - Disposal information	27

#### Introduction

The *Peak Atlas DCA* is an intelligent semiconductor analyser that offers great features together with refreshing simplicity. The *Atlas DCA* brings a world of component data to your fingertips.


#### **Summary Features:**

- Automatic component type identification
  - Bipolar transistors
  - Darlington transistors
  - Enhancement Mode MOSFETs
  - Depletion Mode MOSFETs
  - Junction FETs
  - Low power sensitive Triacs
  - Low power sensitive Thyristors
  - Light Emitting Diodes
  - Bicolour LEDs
  - Diodes
  - Diode networks
- Automatic pinout identification, just connect any way round.
- Special feature identification such as diode protection and resistor shunts.
- Gain measurement for bipolar transistors.
- Leakage current measurement for bipolar transistors.
- Silicon and Germanium detection for bipolar transistors.
- Gate threshold measurement for Enhancement Mode MOSFETs.
- Semiconductor forward voltage measurement for diodes, LEDs and transistor Base-Emitter junctions.
- Automatic and manual power-off.

### **Important Considerations**

#### Please observe the following guidelines:

- This instrument must NEVER be connected to powered equipment/components or equipment/components with <u>any</u> stored energy (e.g. charged capacitors). Failure to comply with this warning may result in personal injury, damage to the equipment under test, damage to the *Atlas DCA* and invalidation of the manufacturer's warranty.
- The *Atlas DCA* is designed to analyse semiconductors that are <u>not</u> in-circuit, otherwise complex circuit effects will result in erroneous measurements.
- Avoid rough treatment or hard knocks.
- This unit is <u>not</u> waterproof.
- Only use a good quality Alkaline battery.





#### **Analysing Components**

The *Atlas DCA* is designed to analyse discrete, unconnected, unpowered components. This ensures that external connections don't influence the measured parameters. The three test probes can be connected to the component any way round. If the component has only two terminals, then any pair of the three test probes can be used.

Peak Atlas DCA model DCA55 Rx.x The *Atlas DCA* will start component analysis when the **on-test** button is pressed.

Depending on the component type, analysis may take a few seconds to complete, after which, the results of the analysis are displayed. Information is displayed a "page" at a time, each page can be displayed by briefly pressing the **scroll-off** button.



The arrow symbol on the display indicates that more pages are available to be viewed.



Although the *Atlas DCA* will switch itself off if left unattended, you can manually switch the unit off by holding down the **scroll-off** button for a couple of seconds.

If the *Atlas DCA* cannot detect any component between any of the test probes, the following message will be displayed:

No component detected

If the component is not a supported component type, a faulty component or a component that is being tested incircuit, the analysis may result in the following message being displayed:

Unknown/Faulty
component

Some components may be faulty due to a shorted junction between a pair of the probes. If this is the case, the following message (or similar) will be displayed:

Short circuit on Green Blue

If all three probes are shorted (or very low resistance) then the following message will be displayed:

Short circuit on Red Green Blue



It is possible that the *Atlas DCA* may detect one or more diode junctions or other component type within an unknown or faulty part. This is because many semiconductors comprise of PN (diode) junctions. Please refer to the section on diodes and diode networks for more information.

#### **Diodes**

The *Atlas DCA* will analyse almost any type of diode. Any pair of the three test clips can be connected to the diode, any way round. If the unit detects a single diode, the following message will be displayed:

Diode or diode junction(s)

RED GREEN BLUE Anod Cath

Forward voltage Vf=0.677V

Test current If=4.300mA Pressing the **scroll-off** button will then display the pinout for the diode.

In this example, the Anode of the diode is connected to the Red test clip and the Cathode is connected to the Green test clip, additionally, the Blue test clip is unconnected. The forward voltage drop is then displayed, this gives an indication of the diode technology. In this example, it is likely that the diode is a silicon diode. A germanium or Schottky diode may yield a forward voltage of about 0.25V. The current at which the diode was tested is also displayed.




Note that the *Atlas DCA* will detect only one diode even if two diodes are connected in series when the third test clip is not connected to the junction between the diodes. The forward voltage drop displayed however will be the voltage across the whole series combination.



The *Atlas DCA* will determine that the diode(s) under test is an LED if the measured forward voltage drop exceeds 1.50V. Please refer to the section on LED analysis for more information.

#### **Diode Networks**

The *Atlas DCA* will intelligently identify popular types of three terminal diode networks. For three terminal devices such as SOT-23 diode networks, the three test clips must all



be connected, any way round. The instrument will identify the type of diode network and then display information regarding each detected diode in sequence. The following types of diode networks are automatically recognised by the *Atlas DCA*:

Common cathode diode network



Both cathodes connected together, such as the BAV70 device.

Common anode diode network



Anodes of each diode are connected together, the BAW56W is an example.

Series diode network



Here, each diode is connected in series. An example is the BAV99.

Following the component identification, the details of each diode in the network will be displayed.

Firstly, the pinout for the diode is displayed, followed by the electrical information, forward voltage drop and the current at which the diode was tested. The value of the test current depends on the measured forward voltage drop of the diode.

Pinout for D1...

RED GREEN BLUE Cath Anod

Forward voltage D1 Uf=0.642U

Following the display of all the details for the first diode, the details of the second diode will then be displayed.

#### **LEDs**

An LED is really just another type of diode, however, the *Atlas DCA* will determine that an LED or LED network has been detected if the measured forward voltage drop is larger than 1.5V. This also enables the *Atlas DCA* to intelligently identify bicolour LEDs, both two-terminal and three-terminal varieties.

LED or diode junction(s)

Like the diode analysis, the pinout, the forward voltage drop and the associated test current is displayed.

RED GREEN BLUE Cath Anod Here, the Cathode (-ve) LED terminal is connected to the Green test clip and the Anode (+ve) LED terminal is connected to the Blue test clip.

Forward voltage Uf=1.936V In this example, a simple green LED yields a forward voltage drop of 1.936V.

Test current If=3.047mA

The test current is dependant on the forward voltage drop of the LED, here the test current is measured as 3.047mA.



Some blue LEDs (and their cousins, white LEDs) require high forward voltages and may not be detected by the *Atlas DCA*.

#### **Bicolour LEDs**

Bicolour LEDs are automatically identified. If your LED has 3 leads then ensure they are all connected, in any order.

A two terminal bicolour LED consists of two LED chips which are connected in inverse parallel within the LED body. Three terminal bicolour LEDs are made with either common anodes or common cathodes.

Two terminal bicolour LED

Three terminal bicolour LED

Here a two terminal LED has been detected.

This message will be displayed if the unit has detected a three terminal LED.



The details of <u>each</u> LED in the package will then be displayed in a similar way to the diode networks detailed earlier.

The pinout for the 1<sup>st</sup> LED is displayed. Remember that this is the pinout for just one of the two LEDs in the package.

Interestingly, the voltage drops for each LED relate to the different colours within the bicolour LED. It may therefore be possible to determine which lead is connected to each colour LED within the device. Red LEDs often have the lowest forward voltage drop, followed by yellow LEDs, green LEDs and finally, blue LEDs.

Pinout for D1...


RED GREEN BLUE Anod Cath

Forward voltage D1 Vf=1.950V

Test current D1 If=3.033mA

### **Bipolar Junction Transistors (BJTs)**

Bipolar Junction Transistors are simply "conventional" transistors, although variants of these do exist such as Darlingtons, diode protected (free-wheeling diode), resistor shunted types and combinations of these types. All of these variations are automatically identified by the *Atlas DCA*.



Bipolar Junction **Transistors** available in two main types, NPN and PNP. In this example, the unit has detected a Silicon PNP transistor.

PMP Silicon Transistor

PNP Germanium Transistor

unit will determine The that transistor is Germanium only if the baseemitter voltage drop is less than 0.55V.



If the device is a Darlington transistor (two BJTs connected together), the unit will display similar a message to this:

NPN Darlineton Transistor



Note that the Atlas DCA will determine that the transistor under test is a Darlington type if the base-emitter voltage drop is greater than 1.00V for devices with a base-emitter shunt resistance of greater than  $60k\Omega$  or if the base-emitter voltage drop is greater than 0.80V for devices with a base-emitter shunt resistance of less than  $60k\Omega$ . The measured base-emitter voltage drop is displayed as detailed later in this section.

Pressing the **scroll-off** button will result in the transistor's pinout being displayed.

Here, the instrument has identified that the Base is connected to the Red test clip, the Collector is connected to the Green test clip and the Emitter is connected to the Blue test clip.

RED GREEN BLUE Base Coll Emit

#### **Transistor Special Features**

Many modern transistors contain additional special features. If the *Atlas DCA* has detected any special features, then the details of these features are displayed next after pressing the **scroll-off** button. If there are no special features detected then the next screen will be the transistor's current gain.

Some transistors, particularly CRT deflection transistors and many large Darlingtons have a protection diode inside their package connected between the collector and emitter.


Diode protection between C-E

The Philips BU505DF is a typical example of a diode protected bipolar transistor. Remember that protection diodes are always internally connected



between the collector and the emitter so that they are normally reverse biased.

For NPN transistors, the anode of the diode is connected to the emitter of the transistor. For PNP transistors, the anode of the diode is connected to the collector of the transistor. Additionally, many Darlingtons and a few non-Darlington transistors also have a resistor shunt network between the base and emitter of the device.



The *Atlas DCA* can detect the resistor shunt if it has a resistance of typically less than  $60k\Omega$ .

The popular Motorola TIP110 NPN Darlington transistor contains internal resistors between the base and emitter.

When the unit detects the presence of a resistive shunt between the base and emitter, the display will show:

Additionally, the *Atlas DCA* will warn you that the accuracy of gain measurement ( $h_{FE}$ ) has been affected by the shunt resistor.

Resistor shunt between B-E


h<sub>FE</sub> not accurate due to B-E res



It is important to note that if a transistor does contain a base-emitter shunt resistor network, any measurements of current gain ( $h_{FE}$ ) will be very low at the test currents used by the *Atlas DCA*. This is due to the resistors providing an additional path for the base current. The readings for gain however can still be used for comparing transistors of a similar type for the purposes of matching or gain band selecting. The *Atlas DCA* will warn you if such a condition arises as illustrated above.

#### **Faulty or Very Low Gain Transistors**

Faulty transistors that exhibit very low gain may cause the *Atlas DCA* to only identify one or more diode junctions within the device. This is because NPN transistors consist of a structure of junctions that behave like a common anode diode network. PNP transistors can appear to be common cathode diode



networks. The common junction represents the base terminal. This is normal

Common anode diode network for situations where the current gain is so low that it is immeasurable at the test currents used by the *Atlas DCA*.



Please note that the equivalent diode pattern may not be correctly identified by the *Atlas DCA* if your transistor is a darlington type or has additional diode(s) in its package (such as a collector-emitter protection diode). This is due to multiple pn junctions that cannot be uniquely analysed.


In some circumstances, the unit may not be able to deduce anything sensible from the device at all, in which case you will see either of these messages:

Unknown/Faulty component No component detected

#### **Current Gain (h**<sub>FE</sub>)

The DC current gain (h<sub>FE</sub>) is displayed after any special transistor features have been displayed.

DC current gain is simply the ratio of the collector current (less leakage) to the base current for a particular operating condition. The *Atlas DCA* measures  $h_{\rm FE}$  at a collector current of 2.50mA and a collector-emitter voltage of between 2V and 3V.



Current eain h<sub>FE</sub>=126

Test current Ic=2.50mA The gain of all transistors can vary considerably with collector current, collector voltage and also temperature. The displayed value for gain therefore may not represent the gain experienced at other collector currents and voltages. This is particularly true for large devices.


Darlington transistors can have very high gain values and more variation of gain will be evident as a result of this.

Additionally, it is quite normal for transistors of the same type to have a wide range of gain values. For this reason, transistor circuits are often designed so that their operation has little dependence on the absolute value of current gain.

The displayed value of gain is very useful however for comparing transistors of a similar type for the purposes of gain matching or fault finding.

#### **Base-Emitter Voltage Drop**

The DC characteristics of the base-emitter junction are displayed, both the base-emitter forward voltage drop and the base current used for the measurement.



B-E Voltage Vbe=0.664V

Test current Ib=4.312mA The forward base-emitter voltage drop can aid in the identification of silicon or germanium devices. Germanium devices can have base-emitter voltages as low as 0.2V, Silicon types exhibit readings of about 0.7V and Darlington transistors can exhibit readings of about 1.2V because of the multiple base-emitter junctions being measured.



Base-Emitter voltage drop measurements can be useful when matching transistors.




Note that the *Atlas DCA* does not perform the base-emitter tests at the same base current as that used for the current gain measurement.

#### **Collector Leakage Current**

The collector current that takes place when no base current is flowing is referred to as *Leakage Current*.

Most modern transistor exhibit extremely low values of leakage current, often less than 1µA, even for very high collector-emitter voltages.



Leakase current Ic=0.170mA Older Germanium types however can suffer from significant collector leakage current, particular at high temperatures (leakage current can be <u>very</u> temperature dependant).


If your transistor is a Silicon type, you should expect to see a leakage current of close to 0.000mA unless the transistor is faulty.




The minimum leakage current that the *Atlas DCA* can measure is typically  $10\mu A$  (0.010mA). For leakage currents higher than  $10\mu A$ , the measurement resolution is typically  $2\mu A$  (0.002mA).

### **Digital Transistors**

Digital transistors aren't really digital, they can act in both a linear or fully on/off mode. They're called "digital transistors" because they can be driven directly by digital outputs without the need for base current limiting resistors.





These parts are most often found in surface mount packages and are becoming more common, particularly in mass produced electronic products.

The presence of the base resistor (and the base-emitter shunt resistor) means that it isn't possible for the *Atlas DCA* to measure the gain of the device, so only the device polarity (NPN/PNP) and pinout is shown.

NPN Disital Transistor

RED GREEN BLUE Emit Base Coll

#### **Enhancement Mode MOSFETs**

MOSFET stands for *Metal Oxide Semiconductor Field Effect Transistor*. Like bipolar transistors, MOSFETs are available in two main types, N-Channel and P-Channel. Most modern MOSFETs are of the Enhancement Mode type, meaning that



Enhancement mode N-Ch MOSFET the bias of the gate-source voltage is always positive (For N-Channel types). The other (rarer) type of MOSFET is the Depletion Mode type which is described in a later section.

MOSFETs of all types are sometimes known as IGFETs, meaning *Insulated Gate Field Effect Transistor*. This term describes a key feature of these devices, an insulated gate region that results in negligible gate current for both positive and negative gate-source voltages (up to the maximum allowed values of course, typically  $\pm 20$ V).

The first screen to be displayed gives information on the type of MOSFET

detected. Pressing **scroll-off** will then result in the pinout of the MOSFET being displayed. The gate, source and drain are each identified.

An important feature of a MOSFET is the gate-source threshold voltage, the gate-source voltage at which conduction between the source and drain starts. The gate threshold is displayed following the pinout information.

RED GREEN BLUE Gate Drn Srce

Gate Threshold Ves=3.47V

Test current Id=2.50mA

The *Atlas DCA* detects that drain-source conduction has started when it reaches 2.50mA.

#### **Depletion Mode MOSFETs**

The fairly rare Depletion Mode MOSFET is very similar to the conventional Junction FET (JFET) except that the gate terminal is insulated from the other two terminals. The input resistance of these devices can typically be greater than  $1000 M\Omega$  for negative and positive gate-source voltages.



Depletion mode N-Ch MOSFET Depletion Mode devices are characterised by the gate-source voltage required to control the drain-source current.


Modern Depletion Mode devices are generally only available in N-Channel varieties and will conduct current between its drain and source terminals even with a zero voltage applied across the gate and the source. The device can only be turned completely off by taking its gate significantly more negative than its source terminal, say –10V. It is this characteristic that makes them so similar to conventional JFETs.

Pressing **scroll-off** will cause the pinout screen to be displayed.

RED GREEN BLUE Drn Gate Srce

## **Junction FETs (JFETs)**

Junction FETs are conventional Field Effect Transistors.



The voltage applied across the gate-source terminals controls current between the drain and source terminals. N-Channel JFETs require a negative voltage on their gate with respect to their source, the more negative the voltage, the less current can flow between the drain and source.

P-Channel
Junction FET

Unlike Depletion Mode MOSFETs, JFETs have no insulation layer on the gate. This means that although the input resistance between the gate and source is


normally very high (greater than  $100\text{M}\Omega$ ), the gate current can rise if the semiconductor junction between the gate and source or between the gate and drain become forward biased. This can happen if the gate voltage becomes about 0.6V higher than either the drain or source terminals for N-Channel devices or 0.6V lower than the drain or source for P-Channel devices.

Drain and Source not identified

RED GREEN BLUE Gate The internal structure of JFETs is essentially symmetrical about the gate terminal, this means that the drain and source terminals are indistinguishable by the *Atlas DCA*. The JFET type and the gate terminal are identified however.

#### **Thyristors (SCRs) and Triacs**

Sensitive low power thyristors (Silicon Controlled Rectifiers - SCRs) and triacs that require gate currents <u>and</u> holding currents of less than 5mA can be identified and analysed with the *Atlas DCA*.



Sensitive or low power thyristor

RED GREEN BLUE Gate Anod Cath Thyristor terminals are the anode, cathode and the gate. The pinout of the thyristor under test will be displayed on the next press of the **scroll-off** button.

Triac terminals are the MT1, MT2 (MT standing for main terminal) and gate. MT1 is the terminal with which gate current is referenced.

Sensitive or low power triac

RED GREEN BLUE MT1 MT2 Gate



- 1. The unit determines that the device under test is a triac by checking the gate trigger quadrants that the device will reliably operate in. Thyristors operate in only one quadrant (positive gate current, positive anode current). Triacs can typically operate in three or four quadrants, hence their use in AC control applications.
- 2. The test currents used by the *Atlas DCA* are kept low (<5mA) to eliminate the possibility of damage to a vast range of component types. Some thyristors and triacs will not operate at low currents and these types cannot be analysed with this instrument. Note also that if only one trigger quadrant of a triac is detected then the unit will conclude that it has found a thyristor. Please see the technical specifications for more details.

### Care of your Atlas DCA

The **Peak Atlas DCA** should provide many years of service if used in accordance with this user guide. Care should be taken not to expose your unit to excessive heat, shock or moisture. Additionally, the battery should be replaced at least every 12 months to reduce the risk of leak damage.

If a low battery warning message appears, immediate replacement of the battery is recommended.

\*\* Warning \*\* Low Battery **O** 

Depending on your variant, replace the battery with a good quality alkaline type that is identified on the rear label.

# Rear Label: AAA (1.5V)

Suitable Types:
Alkaline
AAA, LR03, MN2400, 24A.

## **Rear Label: 23A/MN21 (12V)**

Suitable Types: GP23A, MN21, V23GA, L1028.

The battery can be replaced by placing your *Atlas DCA* face down on a smooth surface and removing the three screws from the rear of the unit.

Following fitting of the new battery, carefully place the rear cover in position, taking care not to trap the test wires.

Do not over-tighten the screws.

Replacement batteries are available directly from Peak Electronic Design Limited and many good electronic/automotive outlets.

#### **Self Test Procedure**

Each time the *Atlas DCA* is powered up, a self test procedure is performed. In addition to a battery voltage test, the unit measures the performance of many internal functions such as the voltage and current sources, amplifiers, analogue to digital converters and test lead multiplexers. If any of these function measurements fall outside tight performance limits, a message will be displayed and the instrument will switch off automatically.

If the problem was caused by a temporary condition on the test clips, such as applying power to the test clips, then simply re-starting the *Atlas DCA* may clear the problem.

If a persistent problem does arise, it is likely that damage has been caused by an external event such as excessive power being applied to the test clips or a large static discharge taking place. If the problem persists, please contact us for further advice, quoting the displayed fault code.



If there is a low battery condition, the automatic self test procedure will not be performed. For this reason, it is highly recommended that the battery is replaced as soon as possible following a "Low Battery" warning.

## **Appendix A - Technical Specifications**

All values are at 20°C unless otherwise specified.

Parameter	Min	Тур	Max	Note
<b>Bipolar Junction Transistors</b>	·	· •	1	'
Measurable gain range (h <sub>FE</sub> )	4		20000	2
Gain resolution		1 h <sub>FE</sub>	2 h <sub>FE</sub>	2,8
Gain accuracy	±3% ±4 h <sub>FE</sub>			2,8
Gain jitter (3 $\sigma$ )		±0.2%		2,9
Gain test voltage V <sub>CEO</sub>	2.0V		3.0V	2
Gain test collector current I <sub>C</sub>		2.50mA ±5%		2
Measurable V <sub>BE</sub> range	0V		1.80V	
V <sub>BE</sub> resolution		1mV	2mV	8
V <sub>BE</sub> accuracy		±2% ±4mV		
Darlington V <sub>BE</sub> range	0.95V	1.00V	1.80V	3
Darlington V <sub>BE</sub> range (shunted)	0.75V	0.80V	1.80V	4
Ge V <sub>BE</sub> range (I <sub>CLEAK</sub> <10μA)	0V		0.50V	
Ge V <sub>BE</sub> range (I <sub>CLEAK</sub> >10μA)	0V		0.55V	
Base-emitter shunt threshold	50kΩ	60kΩ	70kΩ	
Collector leakage test voltage	3.0V	4.0V	5.1V	
Collector leakage range	0.010mA		1.750mA	
Collector leakage resolution		1μA	2μΑ	
Collector leakage accuracy		±2% ±4µA		
Si Acceptable leakage	0mA		0.2mA	6
Ge Acceptable leakage	0mA		1.75mA	6
MOSFETs				
Gate threshold range	0.1V		5.0V	5
Gate threshold accuracy	±2% ±20mV			5
Gate threshold drain current	2.50mA ±5%			
Min. gate-source resistance		8kΩ		
Depletion drain test current	0.5mA		5.5mA	
Diodes/LEDs				
Diode test current			5.0mA	
V <sub>F</sub> resolution		1mV	2mV	
V <sub>F</sub> accuracy	±2% ±4mV			
V <sub>F</sub> for LED identification	1.50V		4.00V	

Continued on next page...

## Appendix A - Technical Specifications continued

All values are at 20°C unless otherwise specified.

Parameter	Min	Тур	Max	Note
JFETs				
Drain-source test current	0.5mA		5.5mA	
SCRs/Triacs				
Gate test current		4.5mA		7
Load test current		5.0mA		
General Specifications				
Peak test current into S/C	-5.5mA		5.5mA	1
Peak test voltage across O/C	-5.1V		5.1V	1
Short circuit threshold	$5\Omega$	10Ω	15Ω	
Analysis duration	1 Sec	3 Secs	6 Secs	
Battery voltage range (AAA)	1.0V	1.5V	1.6V	
Battery voltage range (GP23)	8.0V	12V		
Inactivity power-down period		60 Secs		
Operating temperature range	10°C		40°C	10
	50°F		104°F	10
Battery warning threshold	1.1V (AAA			
Dimensions (body)	10			
	4.1" x 2.8" x 0.8"			

#### **Notes:**

- 1. Between any pair of test clips.
- 2. Collector current of 2.50mA and h<sub>FE</sub>≤2000.
- 3. Resistance across reverse biased base-emitter  $> 60 \text{k}\Omega$ .
- 4. Resistance across reverse biased base-emitter  $< 60 \text{k}\Omega$ .
- 5. Drain-source current of 2.50mA.
- 6.  $V_{CE}$ =4.0V±1.0V. Base automatically tied to emitter with 910k $\Omega$  to reduce pickup.
- 7. Thyristor quadrant I, Triac quadrants I and III.
- 8. BJT with no shunt resistors.
- 9. Tested for Si BJT with  $h_{FE}=1500$ .
- 10. Subject to acceptable LCD visibility.

Please note: Specifications subject to revision.

## **Appendix B – Warranty Information**

**Peak Satisfaction Guarantee** - If for any reason you are not completely satisfied with the *Peak Atlas DCA* within 14 days of purchase you may return the unit to your distributor. You will receive a refund covering the full purchase price if the unit is returned in perfect condition.

**Peak Warranty** - The warranty is valid for 24 months from date of purchase. This warranty covers the cost of repair or replacement due to defects in materials and/or manufacturing faults. The warranty does not cover malfunction or defects caused by:

- a) Operation outside the scope of the user guide.
- b) Unauthorised access or modification (except for battery replacement).
- c) Accidental physical damage or abuse.
- d) Normal wear and tear.

Statutory rights unaffected. Claims must be accompanied by proof of purchase.

## Appendix C – Disposal Information



#### WEEE (Waste of Electrical and Electronic Equipment), Recycling of Electrical and Electronic Products

In 2006 the European Union introduced regulations (WEEE) for the collection and recycling of all waste electrical and electronic equipment. It is no longer permissible to simply throw away electrical and electronic equipment. Instead, these products must enter the recycling process. Each individual EU member state has implemented the WEEE regulations into national law in slightly different ways. Please follow your national law when you want to dispose of any electrical or electronic products. **More details can be obtained from your national WEEE recycling agency.** If in doubt, you may send your Peak Product to us for safe and environmentally responsible disposal.

At Peak Electronic Design Ltd we are committed to continual product development and improvement. The specifications of our products are therefore subject to change without notice.

© 2000-2016 Peak Electronic Design Limited - E&OE
Designed and manufactured in the UK
www.peakelec.co.uk Tel. +44 (0) 1298 70012 Fax. +44 (0) 1298 70046